# **FEATURES:**

- 2 Year Warranty
- Universal 85-264V Input
- One to Four Outputs • 0-70°C Operating Temperature
- Compact 2.5" x 4.25" x 1.2" Size IEC 60601-1 3rd ed. Medical Cert.
  - IEC 62368-1 2nd ed. Certification
  - IEC 60601-1-2 4th ed. EMC
  - Class B Emissions per EN55011/32
  - RoHS Compliant
  - Optional Chassis/Cover





CHASSIS/COVER

**OPEN FRAME** 

### **SAFETY SPECIFICATIONS** UL 62368-1:2014, 2nd Edition Underwriters Laboratories File E137708/E140259 CAN/CSA-C22.2 No. 62368-1-14 AAMI/ANSI ES60601-1:2005/(R) 2012 CAN/CSA-C22.2 No. 60601-1:2014 CB Reports/Certificates (including all IEC 62368-1:2014, 2nd Edition IEC 60601-1:2005/A1:2012 National and Group Deviations) EN 62368-1:2014, 2nd Edition TUV SUD America EN 60601-1:2006/A1:2013 Low Voltage Directive (2014/35/EU of February 2014) RoHS Directive (Recast) (2015/863/EU of March 2015) Electrical Equipment (Safety) Regulations 2016 SI No. 1101 Restriction of the Use of Certain Hazardous Substances in EEE Regulations

2012 SI No. 3032 + 2019 SI No.492
MODEL LISTING

MODEL NO.	OUTPUT 1	OUTPUT 2	OUTPUT 3	OUTPUT 4
SRP-40A-4001	+3.3V/5A	+5V/3A	+12V/0.7A	-12V/0.7A
SRP-40A-4002	+5V/5A	+3.3V/3A	+12V/0.7A	-12V/0.7A
SRP-40A-4003	+5V/5A	-5V/3A	+12V/0.7A	-12V/0.7A
SRP-40A-4004	+5V/5A	-5V/3A	+15V/0.7A	-15V/0.7A
SRP-40A-4005	+5V/5A	+24V/1.5A	+12V/0.7A	-12V/0.7A
SRP-40A-4006	+5V/5A	+24V/1.5A	+15V/0.7A	-15V/0.7A
SRP-40A-4007	+3.3V/3.1A	+5V/1.25A	-24V/.27A	-51.6V/.25A
SRP-40A-3001	+5V/5A	+12V/2A	-12V/0.7A	
SRP-40A-3002	+5V/5A	+15V/2A	-15V/0.7A	
SRP-40A-3003	+24V/1.5A		+15V/0.7A	-15V/0.7A
SRP-40A-3004	+14.5V/1.5A	-14.5V/1.5A	+5V/1A	
SRP-40A-2001	+5V/5A	+24V/1.5A		
SRP-40A-2002	+5V/5A	+12V/3A		
SRP-40A-2003	+5V/5A	-5V/4A		
SRP-40A-2004	+12V/3A	-12V/3A		
SRP-40A-2005	+15V/2.5A	-15V/2A		
SRP-40A-2006	+30V/1.2A		-15V/0.7A	
SRP-40A-2007	+3.3V/5A		+5V/0.7A	
SRP-40A-2008	+6V/5A	+9V/1A		
SRP-40A-2009	+30V/0.5A	-30V/0.5A		
SRP-40A-1001	3.3V/10A			
SRP-40A-1002	5V/8A			
SRP-40A-1003	12V/3.33A			
SRP-40A-1004	15V/2.67A			
SRP-40A-1005	24V/1.67A			
SRP-40A-1006	48V/0.83A			
SRP-40A-1007	9V/4.45A			
SRP-40A-1008	12V/3.33A			

## ORDERING INFORMATION


Consult factory for alternate output configurations. Consult factory for positive, negative or floating Output 2. Specify DC Input when ordering SRP-40A-3003 only. Please specify the following optional features when ordering:

I/O - Isolated Outputs CH - Chassis CO - Cover TS - Terminal Strip

	SKP-4UA
OUT	PUT SPECIFICATIONS
Total Output Power at 50°C <sub>(1)</sub>	40W (33W, 1001)
(See Derating Chart)	
Output Voltage Centering	Output 1: $\pm 0.25\%$ (All outputs
	Output 2: $\pm 5.0\%$ at 50% load)
	Output 3: $\pm 3.0\%$
	Output 4: ± 3.0%
Output Voltage Adjust Range	Output 1: 95 - 105%
Load Regulation	Output 1: 0.5% (10-100% load change)
	Output 2: 5.0% (30-100% load change)
	(2003,4002) 7.0% (30-100% load change)
	Output 3: 0.5% (10-100% load change) Output 4: 0.5% (10-100% load change)
Source Regulation	Outputs 1 – 4: 0.5%
Cross Regulation	Output 2: 5.0% (Output 1
Closs Regulation	Output 3: 0.5% (Output 1 Output 3: 0.5% varied 50-100%)
	Output 4: 0.5%
Output Noise	Outputs 1 - 4: 1.0%
Turn on Overshoot	None
Transient Response	Outputs 1 – 4
Voltage Deviation	5.0%
Recovery Time	2 ms
Load Change	50% to 100%
Output Overvoltage Protection	Output 1: 110% to 150%
Output Overcurrent Protection	Outputs 3 & 4: 110% Min.
Output Overpower Protection	Outputs 1 & 2: 110% Min.
	Outputs cycle on/off, auto recovery
Hold Up Time	10 ms min., 40 W Output, 120V Input
Start Up Time	1 Second
	PUT SPECIFICATIONS
Protection Class	
Source Voltage	85 – 264 Volts AC
Frequency Range	47 – 63 Hz
Source Current	
True RMS	1A at 85V Input
Peak Inrush	30 A
Efficiency	0.66 - 0.80 (Varies by model)
ENVIRON	IMENTAL SPECIFICATIONS
Ambient Operating	0° C to + 70° C
Temperature Range	Derating: See Power Rating Chart
Ambient Storage Temp. Range	- 40° C to + 85° C
	0 1 1 1 1 0 000//00
Temperature Coefficient	Outputs 1 – 4: 0.02%/°C
Temperature Coefficient	3,000m ASL – Operating – Medical 60601-1
Temperature Coefficient  Altitude	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1
Altitude	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating
Altitude	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1
Altitude  GENE  Means of Protection	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating ERAL SPECIFICATIONS
Altitude	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating
Altitude  GENE  Means of Protection  Primary to Secondary  Primary to Ground	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection)
Altitude  GENE  Means of Protection  Primary to Secondary  Primary to Ground  Secondary to Ground	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection)
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength(6, 9)	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength(8, 9) Reinforced Insulation	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(6, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(6, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground <300µA NC, <1000µA SFC <100µA NC, <500µA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB  0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005)
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground <300µA NC, <1000µA SFC <100µA NC, <500µA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED/IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Doperational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Doperational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300µA NC, <1000µA SFC <100µA NC, <500µA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S(IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-6 0.15 to 80MHz, 10V, 80% AM
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz.
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-11 0% Ur, 0.5 cycles, 0-315° 100/240V A
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS  2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-11 0% U <sub>T</sub> , 0.5 cycles, 0-315° 100/240V A 0% U <sub>T</sub> , 1 cycles, 0° 100/240V A
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating <b>ERAL SPECIFICATIONS</b> 2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC  100,000 Hours min., MIL-HDBK-217F, 25° C, GB  0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover <b>S (IEC 60601-1-2:2014, 4<sup>TH</sup> ED./IEC 61000-6-2:2005</b> EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-1 0% Uτ, 0.5 cycles, 0-315° 100/240V A 0% Uτ, 10/12 cycles, 0° 100/240V A
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8,9)</sub> Reinforced Insulation Basic Insulation Doperational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity Voltage Dips	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating <b>ERAL SPECIFICATIONS</b> 2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover <b>S (IEC 60601-1-2:2014, 4<sup>TH</sup> ED./IEC 61000-6-2:2005</b> EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-1 0% UT, 1 cycles, 0° 100/240V A 40% UT, 10/12 cycles, 0° 100/240V A 40% UT, 10/12 cycles, 0° 100/240V B 70% UT, 25/30 cycles, 0° 100/240V B
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity Voltage Dips	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating <b>ERAL SPECIFICATIONS</b> 2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover <b>S (IEC 60601-1-2:2014, 4<sup>TH</sup> ED./IEC 61000-6-2:2005</b> EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-11 0% Ur, 0.5 cycles, 0° 100/240V A 40% Ur, 1 cycles, 0° 100/240V A 70% Ur, 25/30 cycles, 0° 100/240V B 70% Ur, 300 cycles, 0° 100/240V B 70% Ur, 300 cycles, 0° 100/240V B
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMC SPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity Voltage Dips  Voltage Interruptions Radiated Emissions	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating  ERAL SPECIFICATIONS   2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover  S (IEC 60601-1-2:2014, 4 <sup>TH</sup> ED./IEC 61000-6-2:2005 EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-11 0% UT, 0.5 cycles, 0-315° 100/240V A 40% UT, 1 cycles, 0° 100/240V A 70% UT, 25/30 cycles, 0° 100/240V B EN 61000-4-11 0% UT, 35/30 cycles, 0° 100/240V B EN 61000-4-11 0% UT, 35/30 cycles, 0° 100/240V B EN 55011/32 Class B
Means of Protection Primary to Secondary Primary to Ground Secondary to Ground Dielectric Strength <sub>(8, 9)</sub> Reinforced Insulation Basic Insulation Operational Insulation Leakage Current Earth Leakage Touch Current Mean-Time Between Failures Weight  EMCSPECIFICATION Electrostatic Discharge Radiated Electromagnetic Field Electrical Fast Transients/Bursts Surge Immunity Conducted Immunity Magnetic Field Immunity Voltage Dips	3,000m ASL – Operating – Medical 60601-1 5,000m ASL – Operating – ITE/AV – 62368-1 12,192m ASL – Non-Operating <b>ERAL SPECIFICATIONS</b> 2MOPP (Means of Patient Protection) 1MOPP (Means of Patient Protection) Operational Insulation(Consult factory for 1MOPP)  5656 VDC, Primary to Secondary 2121 VDC, Primary to Ground 707 VDC, Secondary to Ground  <300μA NC, <1000μA SFC <100μA NC, <500μA SFC 100,000 Hours min., MIL-HDBK-217F, 25° C, GB 0.49 Lbs. Open Frame 0.85 Lbs. Chassis and Cover <b>S (IEC 60601-1-2:2014, 4<sup>TH</sup> ED /IEC 61000-6-2:2005</b> EN 61000-4-2 ±8KV contact / ±15KV air discharge EN 61000-4-3 80MHz-2.7GHz, 10V/m, 80% AM EN 61000-4-4 ±2 KV, 5KHz/100KHz EN 61000-4-5 ±2 KV line to earth / ±1 KV line to line EN 61000-4-8 30A/m, 60 Hz. EN 61000-4-11 0% U <sub>T</sub> , 0.5 cycles, 0° 100/240V A 0% U <sub>T</sub> , 10/12 cycles, 0° 100/240V B 70% U <sub>T</sub> , 25/30 cycles, 0° 100/240V B 70% U <sub>T</sub> , 25/30 cycles, 0° 100/240V B FN 61000-4-11 0% U <sub>T</sub> , 300 cycles, 0° 100/240V B

All specifications are maximum at 25°C/40W unless otherwise stated, may vary by model and are subject to change without notice.


## SRP-40A SERIES MECHANICAL SPECIFICATIONS



### **APPLICATIONS INFORMATION**

- Each output can deliver its rated current but Total Output Power must not exceed 40W (33W, 1001).
- Generally, adequate cooling is provided when semiconductor case temperatures do not exceed 70°C rise and transformer temperature does not exceed 60°C rise at any specified ambient temperature.
- Sufficient area must be provided around power supply to allow natural movement of air to develop in convection-cooled applications.
- This product is intended for use as a professionally-installed component within information technology, industrial, and medical equipment and is not intended for stand-alone operation.
- A minimum load of 10% is required on Output 1 to ensure proper regulation of remaining outputs.
- This product includes only one fuse in the input circuit. In consideration of Clause 8.11.5 of IEC 60601-1:2005, a second fuse may be required in neutral conductor of the end product
- Peak-to-Peak Output Ripple and Noise is measured directly at the output terminals of the power supply, without the use of the probe ground lead or retractable tip (tip-and-barrel method, 20 MHz bandwidth.
- 8. This product was type-tested and safety-certified using the dielectric strength test voltages listed in Table 6 of IEC 60601-1:2005. In consideration of Clause 8.8.3, care must be taken to insure that the voltage applied to a reinforced insulation does not overstress different types and levels of insulation. Primary and secondary to ground capacitors may need to be disconnected prior to performing a dielectric strength test on the power supply or the end product. It is highly recommended that the DC test voltages listed in DVB.1, Annex DVB of UL 60601-1 1st Edition are not exceeded during a production-line dielectric strength test of the assembled end product. Please consult factory for further information.
- This power supply has been safety-approved and final-tested using a DC dielectric strength test. Please consult factory before performing an AC dielectric strength test.
- Remote-Sense terminals may be used to compensate for cable losses up to 250mV, depending on model. The use of a twisted pair, decoupling capacitors, and an appropriately-rated low-impedance capacitor connected across the load will increase noise immunity.
- Maximum screw penetration into bottom chassis mounting holes is 0.100 inches.
   Maximum screw penetration into side chassis mounting holes is 0.250 inches.
- 12. To comply with emissions specifications, all four mounting hole pads must be electrically connected to a common metal chassis. Chassis/Cover option is recommended. Refer to Operating Instructions for additional information.
- Common RF shielding precautions may need to be taken to assure emissions compliance. Refer to Operating Instructions for additional information.
- Maximum Ambient Temperature is reduced to 40°C with optional Chassis and Cover. See chart below.

# MAXIMUM OUTPUT POWER vs. AMBIENT TEMPERATURE



		CONNECTOR SPECIFICATIONS
P1	AC Input	0.156 friction lock header mates with Tyco 640250-3 or equivalent crimp terminal housing with Tyco 3-640706-1 or equivalent crimp terminal.
P3	DC Output (Single)	0.156 friction lock header mates with Tyco 770849-6 or equivalent crimp terminal housing with Tyco 3-640707-1 or equivalent crimp terminal.
P3	DC Output (Multiple)	0.156 friction lock header mates with Tyco 770849-8 or equivalent crimp terminal housing with Tyco 3-640707-1 or equivalent crimp terminal.
P4,P5	Sense	0.100 friction lock header mates with Molex 22-01-2027 or equivalent crimp terminal housing with Molex 08-50-0114 or equivalent crimp terminal.
G	Ground	0.187 quick disconnect terminal.